

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4340

Name: RITTER 1  
Sterilizer ID: V2178198  
02/28/2024 09:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.4 F 0.0  
2:00 71.6 F 0.2  
4:00 76.4 F 1.0  
6:00 100.4 F 2.9  
8:00 166.9 F 4.8  
10:00 178.9 F 7.7  
12:00 200.3 F 11.0  
14:00 218.8 F 5.1  
16:00 219.3 F 5.1  
18:00 223.5 F 8.2  
20:00 232.1 F 11.5  
22:00 230.5 F 6.2  
24:00 237.9 F 11.1  
26:00 236.0 F 8.1  
28:00 246.6 F 14.0  
30:00 257.2 F 19.6  
32:00 267.0 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.7  
1:00 272.4 F 29.3  
2:00 271.1 F 28.6  
3:00 272.3 F 29.4  
4:00 271.4 F 28.7  
  
Min 271.0 F 28.6  
Max 272.4 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.9 F 0.0  
5:00 229.6 F 0.0  
10:00 211.9 F 0.0  
15:00 209.2 F 0.0  
20:00 216.7 F 0.0  
25:00 230.3 F 0.0  
30:00 230.4 F 0.0  
  
Drying Complete

FILLING: 0:52  
HEATING: 33:35  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:12:11  
CYCLE COMPLETE

02/28/2024 10:57 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_