
Midmark M11 - v1.0.0
Total Cycles: 2818

Name: RITTER 1
Sterilizer ID: V2178198
01/25/2023 10:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.3 F 0.0
2:00 186.1 F 0.0
4:00 193.6 F 4.5
6:00 215.3 F 9.6
8:00 224.2 F 6.9
10:00 235.0 F 10.7
12:00 233.5 F 7.7
14:00 242.4 F 11.2
16:00 246.7 F 13.4
18:00 259.5 F 20.3
20:00 270.7 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.9 F 28.9
2:00 271.9 F 28.2
3:00 272.2 F 28.7
4:00 272.6 F 28.8

Min 271.5 F 28.0
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.1 F 0.0
5:00 214.5 F 0.0
10:00 230.1 F 0.0
15:00 231.3 F 0.0
20:00 229.4 F 0.0
25:00 229.9 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 20:48
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:18
CYCLE COMPLETE

01/25/2023 11:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____