
Midmark M11 - v1.0.5
Total Cycles: 2884

Name: RITTER2
Sterilizer ID: V2313228
09/14/2023 01:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.1 F 0.0
2:00 164.9 F 3.1
4:00 183.7 F 7.0
6:00 202.4 F 11.6
8:00 218.9 F 6.2
10:00 230.9 F 11.1
12:00 232.7 F 8.2
14:00 241.5 F 11.9
16:00 245.3 F 13.0
18:00 257.1 F 19.4
20:00 267.7 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 271.4 F 29.4
2:00 272.9 F 29.8
3:00 271.2 F 29.1
4:00 272.5 F 30.0

Min 270.6 F 28.5
Max 273.3 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.5 F 0.0
5:00 223.2 F 0.0
10:00 196.0 F 0.0
15:00 181.0 F 0.0
20:00 177.6 F 0.0
25:00 174.4 F 0.0
30:00 173.3 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 21:19
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:00:32
CYCLE COMPLETE

09/14/2023 02:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____