
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002641
13:59 31/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:08
STERILIZING 3:18
136.4 C 226kPa 3:18
136.2 C 223kPa 3:48
136.4 C 226kPa 4:18
136.4 C 226kPa 4:48
136.4 C 226kPa 5:18
136.3 C 224kPa 5:48
136.4 C 225kPa 6:18
136.4 C 223kPa 6:48
136.4 C 223kPa 6:48
VENTING 6:48
AIR DRYING 7:37
CYCLE COMPLETE 11:25

Digital Signature #
627F378662FCAE61
