
Midmark M11 - v1.0.0
Total Cycles: 4585

Name: RITTER 1
Sterilizer ID: V2178198
05/27/2024 11:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.9 F 0.0
2:00 121.7 F 1.4
4:00 144.1 F 4.0
6:00 170.1 F 7.5
8:00 190.0 F 11.4
10:00 208.8 F 3.6
12:00 218.3 F 8.1
14:00 230.9 F 12.6
16:00 234.4 F 9.0
18:00 239.6 F 10.5
20:00 244.2 F 12.5
22:00 255.1 F 18.4
24:00 265.9 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 273.2 F 29.5
2:00 271.6 F 28.8
3:00 271.6 F 29.0
4:00 272.6 F 29.4

Min 271.2 F 28.6
Max 273.2 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.0 F 0.0
5:00 240.3 F 0.0
10:00 212.4 F 0.0
15:00 222.8 F 0.0
20:00 230.6 F 0.0
25:00 235.0 F 0.0
30:00 231.1 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 25:52
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:04:20
CYCLE COMPLETE

05/27/2024 12:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____