
Midmark M11 - v1.0.5
Total Cycles: 2669

Name: RITTER2
Sterilizer ID: V2313228
07/19/2023 12:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.8 F 0.0
2:00 195.4 F 4.1
4:00 207.9 F 10.0
6:00 220.0 F 6.6
8:00 233.9 F 11.8
10:00 235.0 F 9.7
12:00 236.8 F 8.9
14:00 249.4 F 15.6
16:00 261.7 F 22.6
18:00 272.8 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.6 F 29.7
2:00 271.6 F 29.2
3:00 272.6 F 29.8
4:00 271.9 F 29.2

Min 271.5 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 223.2 F 0.0
10:00 234.2 F 0.0
15:00 233.8 F 0.0
20:00 226.4 F 0.0
25:00 222.5 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 2:23
HEATING: 18:22
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

07/19/2023 01:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____