
Midmark M11 - v1.0.5
Total Cycles: 3768

Name: RITTER2
Sterilizer ID: V2313228
07/11/2024 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.2 F 0.0
2:00 145.4 F 1.3
4:00 147.4 F 3.7
6:00 174.3 F 7.5
8:00 194.7 F 11.6
10:00 215.0 F 5.4
12:00 226.2 F 9.7
14:00 232.8 F 10.0
16:00 237.5 F 11.0
18:00 234.6 F 8.8
20:00 247.8 F 14.8
22:00 258.9 F 20.9
24:00 268.8 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 271.3 F 29.5
2:00 272.3 F 29.6
3:00 272.7 F 30.0
4:00 272.2 F 30.2

Min 271.0 F 29.0
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 214.7 F 0.0
10:00 214.9 F 0.0
15:00 222.3 F 0.0
20:00 236.8 F 0.0
25:00 229.5 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 25:03
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:03:52
CYCLE COMPLETE

07/11/2024 11:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____