
Midmark M11 - v1.0.5
Total Cycles: 2232

Name: RITTER2
Sterilizer ID: V2313228
04/11/2023 01:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.5 F 0.0
2:00 189.9 F 4.1
4:00 203.0 F 9.2
6:00 223.0 F 7.9
8:00 232.5 F 10.6
10:00 232.1 F 8.2
12:00 240.0 F 10.7
14:00 246.8 F 14.2
16:00 259.5 F 21.2
18:00 270.9 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.5 F 29.7
2:00 271.7 F 29.0
3:00 271.9 F 29.7
4:00 272.4 F 29.5

Min 271.2 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.8 F 0.0
5:00 226.9 F 0.0
10:00 234.9 F 0.0
15:00 229.5 F 0.0
20:00 226.2 F 0.0
25:00 227.3 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 18:45
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

04/11/2023 02:37 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____