
Midmark M11 - v1.0.0
Total Cycles: 2303

Name: RITTER 1
Sterilizer ID: V2178198
09/08/2022 10:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.2 F 0.0
2:00 198.9 F 2.0
4:00 208.8 F 7.0
6:00 222.9 F 13.2
8:00 228.1 F 8.0
10:00 238.0 F 13.3
12:00 238.0 F 10.6
14:00 237.9 F 8.6
16:00 250.4 F 15.6
18:00 262.3 F 22.5
20:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.6 F 29.2
2:00 271.6 F 28.4
3:00 272.5 F 29.4
4:00 272.3 F 28.9

Min 271.5 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.5 F 0.0
5:00 222.6 F 0.0
10:00 230.3 F 0.0
15:00 229.0 F 0.0
20:00 229.1 F 0.0
25:00 230.4 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 20:19
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:42
CYCLE COMPLETE

09/08/2022 11:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____