
Midmark M11 - v1.0.0
Total Cycles: 3920

Name: RITTER 1
Sterilizer ID: V2178198
10/18/2023 01:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.3 F 0.0
2:00 130.2 F 2.0
4:00 158.1 F 5.1
6:00 186.4 F 9.3
8:00 221.0 F 8.3
10:00 226.3 F 7.2
12:00 236.1 F 11.9
14:00 233.8 F 8.0
16:00 245.2 F 12.9
18:00 243.8 F 12.0
20:00 256.1 F 18.4
22:00 267.4 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.0
1:00 272.6 F 28.6
2:00 272.2 F 28.9
3:00 271.8 F 28.6
4:00 271.9 F 28.3

Min 271.0 F 27.9
Max 273.0 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.6 F 0.0
5:00 242.3 F 0.0
10:00 224.4 F 0.0
15:00 207.7 F 0.0
20:00 200.2 F 0.0
25:00 200.6 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 23:23
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:01:55
CYCLE COMPLETE

10/18/2023 02:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____