
Midmark M11 - v1.0.0
Total Cycles: 3875

Name: RITTER 1
Sterilizer ID: V2178198
10/05/2023 03:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.3 F 0.0
2:00 180.2 F 3.3
4:00 205.7 F 7.7
6:00 215.6 F 12.9
8:00 229.3 F 7.7
10:00 240.1 F 13.0
12:00 239.6 F 10.5
14:00 236.8 F 8.9
16:00 251.0 F 15.8
18:00 262.6 F 22.5
20:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.9 F 29.2
2:00 272.2 F 29.2
3:00 271.7 F 28.6
4:00 272.5 F 28.9

Min 271.0 F 28.1
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.5 F 0.0
5:00 201.1 F 0.0
10:00 200.3 F 0.0
15:00 198.9 F 0.0
20:00 196.6 F 0.0
25:00 193.6 F 0.0
30:00 191.2 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 20:07
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 00:59:33
CYCLE COMPLETE

10/05/2023 04:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____