
Midmark M11 - v1.0.0
Total Cycles: 3343

Name: RITTER 1
Sterilizer ID: V2178198
06/06/2023 12:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.5 F 0.0
2:00 179.9 F 0.2
4:00 204.8 F 5.1
6:00 226.4 F 10.2
8:00 231.6 F 7.0
10:00 237.4 F 11.3
12:00 235.7 F 8.7
14:00 240.1 F 10.1
16:00 247.4 F 14.0
18:00 259.6 F 20.6
20:00 270.5 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.3
1:00 272.6 F 29.1
2:00 271.6 F 28.2
3:00 272.7 F 29.1
4:00 271.9 F 28.5

Min 271.4 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 218.9 F 0.0
10:00 227.7 F 0.0
15:00 235.6 F 0.0
20:00 228.8 F 0.0
25:00 226.8 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 20:41
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

06/06/2023 01:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____