
Midmark M11 - v1.0.0
Total Cycles: 5430

Name: RITTER 1
Sterilizer ID: V2178198
03/21/2025 11:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.5 F 0.0
2:00 154.6 F 2.6
4:00 160.3 F 5.7
6:00 184.5 F 9.2
8:00 208.0 F 12.3
10:00 221.0 F 6.1
12:00 227.7 F 10.2
14:00 234.0 F 8.3
16:00 239.2 F 11.2
18:00 236.7 F 8.9
20:00 249.1 F 15.0
22:00 259.8 F 20.9
24:00 269.8 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.4 F 29.7
2:00 272.7 F 29.4
3:00 271.6 F 28.9
4:00 271.5 F 29.3

Min 271.2 F 28.8
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.0 F 0.0
5:00 224.3 F 0.0
10:00 228.0 F 0.0
15:00 230.2 F 0.0
20:00 227.1 F 0.0
25:00 231.0 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 25:54
STERILIZING: 4:00
VENTING: 3:60
DRYING: 30:00
TOTAL CYCLE: 01:04:47
CYCLE COMPLETE

03/21/2025 12:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____