
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002417
10:12 13/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:37
STERILIZING 3:41
136.0 C 220kPa 3:41
136.1 C 221kPa 4:11
136.4 C 226kPa 4:41
136.2 C 224kPa 5:11
136.3 C 225kPa 5:41
136.3 C 224kPa 6:11
136.3 C 224kPa 6:41
136.2 C 223kPa 7:11
136.2 C 223kPa 7:11
VENTING 7:11
AIR DRYING 8:05
CYCLE COMPLETE 8:31

Digital Signature #
DDD35C5DB6160040
