
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003816
9:19 15/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:01
STERILIZING 3:08
136.4 C 221kPa 3:08
136.3 C 221kPa 3:38
136.4 C 223kPa 4:08
136.4 C 222kPa 4:38
136.4 C 222kPa 5:08
136.4 C 223kPa 5:38
136.0 C 220kPa 6:08
136.3 C 223kPa 6:38
136.3 C 223kPa 6:38
VENTING 6:38
AIR DRYING 7:24
CYCLE COMPLETE 10:00

Digital Signature #
1094DD90833CCE6D
