
Midmark M11 - v1.0.0
Total Cycles: 4892

Name: RITTER 1
Sterilizer ID: V2178198
09/11/2024 01:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.0 F 0.0
2:00 145.2 F 0.2
4:00 172.1 F 0.8
6:00 196.5 F 4.5
8:00 210.6 F 8.5
10:00 224.2 F 13.2
12:00 228.9 F 7.9
14:00 240.5 F 12.8
16:00 239.3 F 10.3
18:00 236.1 F 8.3
20:00 250.6 F 15.3
22:00 262.3 F 21.9
24:00 272.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.8 F 28.7
2:00 271.7 F 27.9
3:00 272.2 F 28.6
4:00 272.6 F 28.5

Min 271.4 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 206.5 F 0.0
10:00 205.7 F 0.0
15:00 215.6 F 0.0
20:00 232.6 F 0.0
25:00 230.2 F 0.0
30:00 230.3 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 24:23
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:02:41
CYCLE COMPLETE

09/11/2024 02:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____