
Midmark M11 - v1.0.0
Total Cycles: 4792

Name: RITTER 1
Sterilizer ID: V2178198
08/07/2024 05:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.7 F 0.0
2:00 172.7 F 2.9
4:00 191.8 F 7.4
6:00 220.0 F 12.6
8:00 226.5 F 7.0
10:00 237.9 F 12.2
12:00 236.2 F 9.2
14:00 237.9 F 9.2
16:00 248.4 F 14.4
18:00 260.9 F 21.3
20:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.2 F 28.7
2:00 271.5 F 28.6
3:00 272.3 F 28.5
4:00 271.4 F 28.6

Min 270.9 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.1 F 0.0
5:00 194.9 F 0.0
10:00 195.3 F 0.0
15:00 192.7 F 0.0
20:00 193.4 F 0.0
25:00 196.2 F 0.0
30:00 199.8 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

08/07/2024 06:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____