

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001585  
14:53 11/15/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:49  
PRESSURIZING 5:56  
STERILIZING 7:42  
136.0 C 220kPa 7:42  
136.3 C 222kPa 8:12  
136.4 C 223kPa 8:42  
136.5 C 220kPa 9:12  
136.4 C 222kPa 9:42  
136.2 C 221kPa 10:12  
136.4 C 221kPa 10:42  
136.4 C 221kPa 11:12  
VENTING 11:12  
AIR DRYING 12:16  
CYCLE COMPLETE 20:11  
-----

Digital Signature #  
EBE3275E58099EA2  
-----