
Midmark M11 - v1.0.0
Total Cycles: 927

Name: RITTER 1
Sterilizer ID: V2178198
06/23/2021 08:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.1 F 0.0
2:00 87.6 F 0.0
4:00 99.4 F 0.0
6:00 109.6 F 0.0
8:00 146.8 F 0.1
10:00 184.7 F 0.2
12:00 209.6 F 1.6
14:00 222.7 F 4.6
16:00 227.9 F 7.2
18:00 232.9 F 10.5
20:00 234.1 F 9.3
22:00 239.1 F 10.7
24:00 235.0 F 7.8
26:00 244.8 F 12.0
28:00 245.4 F 12.5
30:00 257.1 F 18.7
32:00 267.4 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 27.9
1:00 272.6 F 28.5
2:00 271.2 F 27.9
3:00 272.7 F 28.5
4:00 271.4 F 27.8

Min 271.1 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 235.4 F 0.0
10:00 218.4 F 0.0
15:00 220.2 F 0.0
20:00 217.1 F 0.0
25:00 217.7 F 0.0
30:00 212.8 F 0.0

Drying Complete

FILLING: 1:08
HEATING: 33:29
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:11:45
CYCLE COMPLETE

06/23/2021 09:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____