

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 008429  
7:54 20/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:50  
PRESSURIZING 1:55  
STERILIZING 2:59  
136.4 C 221kPa 2:59  
136.4 C 223kPa 3:29  
136.3 C 219kPa 3:59  
136.4 C 218kPa 4:29  
136.2 C 218kPa 4:59  
136.4 C 221kPa 5:29  
136.2 C 216kPa 5:59  
136.3 C 221kPa 6:29  
136.3 C 221kPa 6:29  
VENTING 6:29  
AIR DRYING 7:21  
CYCLE COMPLETE 37:21  
-----

Digital Signature #  
826274043F9AED3D  
-----