
Midmark M11 - v1.0.5
Total Cycles: 2146

Name: RITTER2
Sterilizer ID: V2313228
03/14/2023 10:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.2 F 0.0
2:00 188.0 F 3.8
4:00 199.6 F 9.2
6:00 221.3 F 7.2
8:00 232.3 F 11.1
10:00 234.5 F 9.3
12:00 237.1 F 9.4
14:00 249.8 F 15.7
16:00 262.4 F 23.0
18:00 272.3 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.3 F 29.5
2:00 271.9 F 29.6
3:00 272.2 F 29.4
4:00 271.7 F 29.5

Min 271.4 F 29.0
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.7 F 0.0
5:00 223.9 F 0.0
10:00 225.6 F 0.0
15:00 229.2 F 0.0
20:00 226.2 F 0.0
25:00 223.5 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 18:11
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:22
CYCLE COMPLETE

03/14/2023 11:44 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____