
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001501
10:36 10/14/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:57
PRESSURIZING 6:39
STERILIZING 7:57
136.0 C 220kPa 7:57
136.5 C 222kPa 8:27
136.4 C 222kPa 8:57
136.3 C 222kPa 9:27
136.4 C 220kPa 9:57
136.4 C 222kPa 10:27
136.4 C 222kPa 10:57
136.0 C 220kPa 11:27
VENTING 11:27
AIR DRYING 12:36
CYCLE COMPLETE 17:01

Digital Signature #
60F8F922A9CFEE65
