
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008336
8:38 24/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:15
STERILIZING 4:19
136.4 C 221kPa 4:19
136.2 C 221kPa 4:49
135.8 C 220kPa 5:19
136.4 C 219kPa 5:49
136.3 C 219kPa 6:19
136.4 C 223kPa 6:49
136.1 C 218kPa 7:19
136.4 C 219kPa 7:49
136.4 C 219kPa 7:49
VENTING 7:49
AIR DRYING 8:36
CYCLE COMPLETE 9:03

Digital Signature #
8C1893C2809DEB4E
