
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004944
12:53 10/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 2:59
STERILIZING 4:05
136.5 C 222kPa 4:05
136.2 C 224kPa 4:35
136.4 C 224kPa 5:05
136.4 C 225kPa 5:35
136.2 C 218kPa 6:05
136.3 C 222kPa 6:35
136.4 C 224kPa 7:05
136.2 C 221kPa 7:35
136.2 C 221kPa 7:35
VENTING 7:35
AIR DRYING 8:26
CYCLE COMPLETE 15:11

Digital Signature #
BBD30E755474FA04
