
Midmark M11 - v1.0.0
Total Cycles: 4447

Name: RITTER 1
Sterilizer ID: V2178198
04/11/2024 11:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.1 F 0.0
2:00 136.2 F 1.5
4:00 146.3 F 4.5
6:00 210.4 F 7.8
8:00 197.9 F 12.3
10:00 220.3 F 4.5
12:00 225.1 F 8.4
14:00 235.5 F 12.7
16:00 235.3 F 8.9
18:00 241.6 F 11.5
20:00 244.6 F 12.5
22:00 256.1 F 18.6
24:00 266.9 F 25.1
26:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.0 F 28.7
2:00 272.8 F 29.0
3:00 272.3 F 29.2
4:00 271.5 F 28.6

Min 271.5 F 28.5
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.4 F 0.0
5:00 227.5 F 0.0
10:00 230.6 F 0.0
15:00 227.9 F 0.0
20:00 223.7 F 0.0
25:00 230.8 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 26:13
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:04:49
CYCLE COMPLETE

04/11/2024 12:12 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____