
Midmark M11 - v1.0.5
Total Cycles: 3921

Name: RITTER2
Sterilizer ID: V2313228
09/10/2024 08:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 71.2 F 0.0
4:00 77.1 F 0.0
6:00 101.0 F 0.0
8:00 144.2 F 0.2
10:00 177.9 F 0.3
12:00 202.5 F 2.9
14:00 209.3 F 6.0
16:00 217.9 F 9.4
18:00 230.5 F 13.2
20:00 231.7 F 8.6
22:00 242.2 F 13.4
24:00 240.9 F 11.3
26:00 238.6 F 10.0
28:00 251.8 F 16.4
30:00 262.6 F 22.7
32:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.3 F 28.8
2:00 272.5 F 29.3
3:00 271.6 F 28.6
4:00 272.4 F 29.0

Min 271.5 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.0 F 0.0
5:00 207.7 F 0.0
10:00 200.8 F 0.0
15:00 196.1 F 0.0
20:00 191.2 F 0.0
25:00 189.0 F 0.0
30:00 188.5 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 32:21
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:10:29
CYCLE COMPLETE

09/10/2024 09:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____