
Midmark M11 - v1.0.5
Total Cycles: 1826

Name: RITTER2
Sterilizer ID: V2313228
12/08/2022 01:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.5 F 0.0
2:00 189.3 F 4.0
4:00 198.7 F 9.2
6:00 218.5 F 7.1
8:00 230.7 F 10.4
10:00 232.6 F 7.8
12:00 240.5 F 10.7
14:00 247.7 F 14.4
16:00 260.3 F 21.5
18:00 271.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.4 F 29.2
2:00 272.3 F 29.7
3:00 272.1 F 29.0
4:00 272.4 F 29.7

Min 271.6 F 28.9
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.9 F 0.0
5:00 228.8 F 0.0
10:00 249.2 F 0.0
15:00 227.5 F 0.0
20:00 227.1 F 0.0
25:00 236.8 F 0.0
30:00 220.4 F 0.0
Drying Complete

FILLING: 2:00
HEATING: 18:33
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:41
CYCLE COMPLETE

12/08/2022 02:09 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____