
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 000592
15:24 19/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:49
PRESSURIZING 2:56
STERILIZING 3:59
136.0 C 227kPa 3:59
136.1 C 231kPa 4:29
136.2 C 230kPa 4:59
136.4 C 233kPa 5:29
136.3 C 231kPa 5:59
136.2 C 229kPa 6:29
136.4 C 232kPa 6:59
136.4 C 233kPa 7:29
136.4 C 233kPa 7:29
VENTING 7:29
AIR DRYING 8:21
CYCLE COMPLETE 23:21

Digital Signature #
54AE760445AD0061
