
Midmark M11 - v1.0.5
Total Cycles: 1052

Name: RITTER2
Sterilizer ID: V2313228
05/02/2022 08:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.8 F 0.0
2:00 156.0 F 2.1
4:00 159.6 F 5.1
6:00 185.2 F 9.7
8:00 218.2 F 6.3
10:00 227.1 F 9.1
12:00 233.6 F 9.9
14:00 240.8 F 11.7
16:00 238.9 F 10.1
18:00 251.8 F 16.7
20:00 263.4 F 23.5
22:00 271.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 271.5 F 29.3
2:00 272.4 F 29.6
3:00 272.0 F 29.9
4:00 271.5 F 29.7

Min 271.1 F 29.3
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 233.3 F 0.0
10:00 226.5 F 0.0
15:00 219.3 F 0.0
20:00 196.7 F 0.0
25:00 199.2 F 0.0
30:00 192.0 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 22:49
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:28
CYCLE COMPLETE

05/02/2022 09:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____