
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001719
13:53 21/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:36
STERILIZING 3:43
136.4 C 224kPa 3:43
136.3 C 225kPa 4:13
136.4 C 224kPa 4:43
136.1 C 222kPa 5:13
136.4 C 226kPa 5:43
136.2 C 224kPa 6:13
136.4 C 226kPa 6:43
136.4 C 224kPa 7:13
136.4 C 224kPa 7:13
VENTING 7:13
AIR DRYING 8:07
CYCLE COMPLETE 38:07

Digital Signature #
B98DF33171289051
