

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3271

Name: RITTER2  
Sterilizer ID: V2313228  
01/03/2024 01:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.6 F 0.0  
2:00 148.6 F 1.5  
4:00 162.6 F 4.7  
6:00 190.0 F 8.5  
8:00 203.0 F 12.9  
10:00 218.9 F 6.5  
12:00 230.4 F 11.0  
14:00 231.2 F 7.9  
16:00 242.6 F 12.8  
18:00 244.1 F 12.7  
20:00 256.0 F 19.0  
22:00 266.5 F 25.6  
24:00 272.7 F 30.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.6  
1:00 272.5 F 29.8  
2:00 272.7 F 30.2  
3:00 272.0 F 29.4  
4:00 272.7 F 30.0

Min 271.8 F 29.4  
Max 272.9 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.5 F 0.0  
5:00 214.5 F 0.0  
10:00 212.8 F 0.0  
15:00 203.9 F 0.0  
20:00 201.3 F 0.0  
25:00 199.8 F 0.0  
30:00 199.4 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 24:24  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:03:09  
CYCLE COMPLETE

01/03/2024 02:35 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_