
Midmark M11 - v1.0.5
Total Cycles: 79

Name: RITTER2
Sterilizer ID: V2313228
06/08/2021 02:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.9 F 0.0
2:00 163.1 F 2.6
4:00 173.7 F 6.1
6:00 194.7 F 10.0
8:00 207.6 F 7.0
10:00 221.3 F 8.4
12:00 233.7 F 13.1
14:00 237.6 F 10.6
16:00 236.6 F 9.1
18:00 249.3 F 15.3
20:00 260.1 F 21.4
22:00 270.2 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.3 F 29.6
2:00 271.3 F 28.9
3:00 272.4 F 29.7
4:00 271.8 F 29.2

Min 271.1 F 28.9
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.4 F 0.0
5:00 231.0 F 0.0
10:00 212.0 F 0.0
15:00 208.2 F 0.0
20:00 212.7 F 0.0
25:00 207.5 F 0.0
30:00 207.2 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 22:58
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:01:60
CYCLE COMPLETE

06/08/2021 03:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____