
Midmark M11 - v1.0.5
Total Cycles: 168

Name: RITTER2
Sterilizer ID: V2313228
07/06/2021 01:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 90.4 F 0.0
2:00 94.4 F 0.7
4:00 107.1 F 2.5
6:00 138.3 F 5.5
8:00 169.0 F 9.1
10:00 191.0 F 13.1
12:00 209.1 F 5.2
14:00 221.2 F 9.1
16:00 232.7 F 13.5
18:00 235.2 F 10.3
20:00 235.6 F 9.0
22:00 246.7 F 14.4
24:00 258.1 F 20.6
26:00 268.2 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 271.5 F 29.4
2:00 271.6 F 29.3
3:00 271.8 F 29.5
4:00 272.1 F 29.7

Min 271.4 F 29.3
Max 272.3 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.3 F 0.0
5:00 203.5 F 0.0
10:00 217.3 F 0.0
15:00 217.0 F 0.0
20:00 215.2 F 0.0
25:00 214.1 F 0.0
30:00 207.5 F 0.0

Drying Complete

FILLING: 0:55
HEATING: 27:09
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:05:13
CYCLE COMPLETE

07/06/2021 02:32 PM
ACCEPTED: X
REJECTED:
Operator: U02