
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000900
16:05 02/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:15
STERILIZING 4:18
136.0 C 228kPa 4:18
136.4 C 233kPa 4:48
136.4 C 233kPa 5:18
136.4 C 233kPa 5:48
136.3 C 232kPa 6:18
136.4 C 232kPa 6:48
136.2 C 232kPa 7:18
136.2 C 231kPa 7:48
136.2 C 231kPa 7:48
VENTING 7:48
AIR DRYING 8:43
CYCLE COMPLETE 16:01

Digital Signature #
DFCFA265C93DC576
