
Midmark M11 - v1.0.0
Total Cycles: 5355

Name: RITTER 1
Sterilizer ID: V2178198
02/21/2025 12:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.9 F 0.0
2:00 123.3 F 1.5
4:00 146.2 F 4.2
6:00 169.9 F 7.5
8:00 191.1 F 11.2
10:00 209.8 F 3.7
12:00 217.8 F 7.1
14:00 231.6 F 11.5
16:00 230.0 F 7.0
18:00 242.0 F 12.3
20:00 239.7 F 10.2
22:00 251.6 F 16.1
24:00 262.1 F 22.3
26:00 272.1 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.5 F 29.3
2:00 272.5 F 28.9
3:00 271.4 F 28.4
4:00 271.5 F 28.8

Min 271.2 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 201.6 F 0.0
10:00 192.4 F 0.0
15:00 185.0 F 0.0
20:00 183.5 F 0.0
25:00 183.2 F 0.0
30:00 184.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 26:34
STERILIZING: 4:00
VENTING: 3:56
DRYING: 30:00
TOTAL CYCLE: 01:05:20
CYCLE COMPLETE

02/21/2025 01:40 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____