
Midmark M11 - v1.0.5
Total Cycles: 2083

Name: RITTER2
Sterilizer ID: V2313228
02/23/2023 09:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.4 F 0.0
2:00 173.3 F 0.6
4:00 178.3 F 3.4
6:00 210.6 F 7.9
8:00 227.9 F 13.4
10:00 231.5 F 9.5
12:00 235.5 F 8.8
14:00 244.1 F 13.4
16:00 243.4 F 12.4
18:00 256.6 F 19.3
20:00 268.3 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.8
1:00 272.5 F 29.9
2:00 272.4 F 29.4
3:00 272.1 F 29.8
4:00 273.0 F 29.7

Min 271.0 F 28.6
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.5 F 0.0
5:00 218.9 F 0.0
10:00 204.7 F 0.0
15:00 223.7 F 0.0
20:00 239.8 F 0.0
25:00 235.1 F 0.0
30:00 212.2 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 21:15
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:23
CYCLE COMPLETE

02/23/2023 10:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____