

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4249

Name: RITTER 1  
Sterilizer ID: V2178198  
01/30/2024 04:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 86.7 F 0.0  
2:00 97.7 F 0.5  
4:00 100.9 F 2.0  
6:00 125.1 F 4.4  
8:00 153.5 F 7.3  
10:00 175.8 F 10.7  
12:00 206.0 F 5.1  
14:00 212.1 F 5.5  
16:00 221.8 F 9.0  
18:00 230.3 F 12.8  
20:00 232.2 F 8.5  
22:00 243.4 F 13.4  
24:00 241.8 F 11.2  
26:00 252.8 F 16.8  
28:00 263.1 F 22.8  
30:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.7  
1:00 272.2 F 28.8  
2:00 272.7 F 29.3  
3:00 271.9 F 28.9  
4:00 272.1 F 28.8

Min 271.7 F 28.6  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.4 F 0.0  
5:00 205.6 F 0.0  
10:00 212.1 F 0.0  
15:00 212.3 F 0.0  
20:00 224.3 F 0.0  
25:00 230.1 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 31:02  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:09:38  
CYCLE COMPLETE

01/30/2024 05:52 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_