
Midmark M11 - v1.0.5
Total Cycles: 2639

Name: RITTER2
Sterilizer ID: V2313228
07/12/2023 03:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.8 F 0.0
2:00 109.4 F 0.0
4:00 134.4 F 0.6
6:00 169.4 F 4.8
8:00 193.3 F 9.7
10:00 209.9 F 6.4
12:00 224.4 F 7.9
14:00 236.7 F 12.8
16:00 237.0 F 9.9
18:00 236.8 F 8.8
20:00 248.8 F 14.9
22:00 260.7 F 21.6
24:00 271.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.8 F 29.3
2:00 271.5 F 28.4
3:00 272.7 F 29.5
4:00 272.4 F 28.9

Min 271.4 F 28.4
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 168.5 F 0.0
5:00 210.4 F 0.0
10:00 224.7 F 0.0
15:00 228.4 F 0.0
20:00 225.8 F 0.0
25:00 222.6 F 0.0
30:00 219.2 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 24:37
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 01:02:30
CYCLE COMPLETE

07/12/2023 04:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____