

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2582

Name: RITTER 1  
Sterilizer ID: V2178198  
11/21/2022 08:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.0 F 0.0  
2:00 76.5 F 0.0  
4:00 109.4 F 0.1  
6:00 141.9 F 1.9  
8:00 151.2 F 4.5  
10:00 173.1 F 7.7  
12:00 193.3 F 11.5  
14:00 207.6 F 2.9  
16:00 219.4 F 7.1  
18:00 230.5 F 11.0  
20:00 230.4 F 6.8  
22:00 240.9 F 11.7  
24:00 237.6 F 9.3  
26:00 250.6 F 15.6  
28:00 261.3 F 21.8  
30:00 271.7 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.8  
1:00 272.5 F 29.1  
2:00 272.5 F 29.5  
3:00 271.7 F 28.7  
4:00 272.7 F 29.3

Min 271.6 F 28.7  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.4 F 0.0  
5:00 217.4 F 0.0  
10:00 222.8 F 0.0  
15:00 222.0 F 0.0  
20:00 232.6 F 0.0  
25:00 223.6 F 0.0  
30:00 219.8 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 31:18  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:09:20  
CYCLE COMPLETE

11/21/2022 09:50 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_