

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4248

Name: RITTER2  
Sterilizer ID: V2313228  
01/08/2025 01:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 118.8 F 0.0  
2:00 124.0 F 0.0  
4:00 127.7 F 0.0  
6:00 163.4 F 0.3  
8:00 192.3 F 1.7  
10:00 209.3 F 5.3  
12:00 222.6 F 8.9  
14:00 229.8 F 13.1  
16:00 232.4 F 9.2  
18:00 239.7 F 11.2  
20:00 242.9 F 12.6  
22:00 242.2 F 11.7  
24:00 254.4 F 18.0  
26:00 265.2 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.4 F 29.4  
2:00 271.6 F 29.3  
3:00 272.3 F 29.2  
4:00 271.6 F 29.3

Min 271.3 F 28.8  
Max 272.5 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.8 F 0.0  
5:00 235.1 F 0.0  
10:00 209.3 F 0.0  
15:00 222.9 F 0.0  
20:00 202.8 F 0.0  
25:00 181.4 F 0.0  
30:00 202.7 F 0.0

Drying Complete

FILLING: 0:58  
HEATING: 27:54  
STERILIZING: 4:00  
VENTING: 3:46  
DRYING: 30:00  
TOTAL CYCLE: 01:06:38  
CYCLE COMPLETE

01/08/2025 02:20 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_