

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3237

Name: RITTER 1  
Sterilizer ID: V2178198  
05/16/2023 07:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 77.1 F 0.0  
2:00 78.2 F 0.0  
4:00 86.1 F 0.0  
6:00 114.7 F 0.0  
8:00 150.4 F 1.7  
10:00 172.8 F 4.9  
12:00 191.4 F 8.4  
14:00 205.1 F 12.0  
16:00 215.3 F 5.0  
18:00 225.6 F 9.0  
20:00 235.1 F 13.2  
22:00 236.7 F 9.8  
24:00 237.5 F 9.2  
26:00 246.8 F 13.7  
28:00 258.4 F 20.0  
30:00 269.1 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.4  
1:00 272.0 F 28.8  
2:00 271.4 F 28.6  
3:00 272.0 F 28.8  
4:00 271.3 F 28.6  
  
Min 271.1 F 28.2  
Max 272.4 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.7 F 0.0  
5:00 197.8 F 0.0  
10:00 208.8 F 0.0  
15:00 219.6 F 0.0  
20:00 227.0 F 0.0  
25:00 231.0 F 0.0  
30:00 229.6 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 31:06  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:09:10  
CYCLE COMPLETE

05/16/2023 08:23 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_