
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001291
12:39 12/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:44
PRESSURIZING 3:48
STERILIZING 5:16
136.0 C 220kPa 5:16
136.4 C 224kPa 5:46
136.3 C 223kPa 6:16
136.3 C 223kPa 6:46
136.4 C 225kPa 7:16
136.4 C 224kPa 7:46
136.0 C 222kPa 8:16
136.4 C 224kPa 8:46
136.4 C 224kPa 8:46
VENTING 8:46
AIR DRYING 9:45
CYCLE COMPLETE 24:45

Digital Signature #
37500623F70C8207
