

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001816  
11:57 08/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:39  
PRESSURIZING 2:43  
STERILIZING 3:49  
136.4 C 224kPa 3:49  
136.4 C 225kPa 4:19  
136.4 C 223kPa 4:49  
136.4 C 225kPa 5:19  
136.5 C 222kPa 5:49  
136.4 C 225kPa 6:19  
136.4 C 225kPa 6:49  
136.2 C 223kPa 7:19  
136.2 C 223kPa 7:19  
VENTING 7:19  
AIR DRYING 8:10  
CYCLE COMPLETE 8:33  
-----

Digital Signature #  
5C44C7D02ABD262C  
-----