
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001538
9:08 24/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:19
STERILIZING 4:30
136.0 C 220kPa 4:30
136.0 C 221kPa 5:00
136.1 C 223kPa 5:30
136.2 C 222kPa 6:00
136.4 C 224kPa 6:30
136.4 C 224kPa 7:00
136.4 C 223kPa 7:30
136.4 C 224kPa 8:00
136.4 C 224kPa 8:00
VENTING 8:00
AIR DRYING 8:56
CYCLE COMPLETE 9:15

Digital Signature #
67FD01BD80F7FA4B
