

-----  
Mississauga, ON  
Canada  
-----

STATIM 5000 S501R751  
SN 511020J00004  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000837  
16:30 31/03/2022  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:40  
PRESSURIZING 6:23  
STERILIZING 8:06  
136.2 C 222kPa 8:06  
136.4 C 222kPa 8:36  
136.4 C 222kPa 9:06  
136.2 C 222kPa 9:36  
136.3 C 223kPa 10:06  
136.3 C 220kPa 10:36  
136.4 C 222kPa 11:06  
136.0 C 220kPa 11:36  
136.0 C 220kPa 11:36  
VENTING 11:36  
AIR DRYING 12:45  
CYCLE COMPLETE 29:49  
-----

Digital Signature #  
5E9D5682316A26A1  
-----