
Midmark M11 - v1.0.5
Total Cycles: 699

Name: RITTER2
Sterilizer ID: V2313228
01/17/2022 08:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.5 F 0.0
2:00 79.6 F 0.3
4:00 109.8 F 1.1
6:00 136.3 F 3.2
8:00 166.9 F 6.7
10:00 173.5 F 10.8
12:00 202.7 F 3.5
14:00 215.0 F 7.1
16:00 224.7 F 11.4
18:00 228.0 F 7.5
20:00 240.5 F 12.8
22:00 239.4 F 10.9
24:00 251.8 F 17.2
26:00 263.1 F 23.8
28:00 271.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.4
1:00 271.8 F 29.6
2:00 272.4 F 30.1
3:00 271.7 F 29.8
4:00 271.8 F 29.4

Min 271.1 F 29.3
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.4 F 0.0
5:00 212.7 F 0.0
10:00 203.8 F 0.0
15:00 200.0 F 0.0
20:00 198.2 F 0.0
25:00 234.1 F 0.0
30:00 216.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 28:09
STERILIZING: 4:00
VENTING: 2:54
DRYING: 30:00
TOTAL CYCLE: 01:06:02
CYCLE COMPLETE

01/17/2022 09:41 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____