

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5456

Name: RITTER 1  
Sterilizer ID: V2178198  
03/27/2025 01:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.1 F 0.0  
2:00 159.0 F 3.0  
4:00 174.1 F 6.8  
6:00 192.4 F 10.9  
8:00 216.3 F 4.5  
10:00 226.6 F 8.6  
12:00 235.3 F 13.3  
14:00 237.5 F 9.6  
16:00 239.5 F 9.6  
18:00 246.5 F 13.4  
20:00 257.8 F 19.5  
22:00 268.2 F 26.0  
24:00 271.7 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.8  
1:00 272.7 F 29.0  
2:00 272.5 F 29.4  
3:00 271.6 F 28.7  
4:00 272.6 F 29.3

Min 271.3 F 28.7  
Max 273.0 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.0 F 0.0  
5:00 204.0 F 0.0  
10:00 215.1 F 0.0  
15:00 222.4 F 0.0  
20:00 248.7 F 0.0  
25:00 240.2 F 0.0  
30:00 231.5 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 24:03  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:02:52  
CYCLE COMPLETE

03/27/2025 02:25 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_