
Midmark M11 - v1.0.5
Total Cycles: 1275

Name: RITTER2
Sterilizer ID: V2313228
07/08/2022 01:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.6 F 0.0
2:00 132.2 F 1.7
4:00 155.2 F 5.1
6:00 183.6 F 9.6
8:00 213.0 F 6.4
10:00 225.3 F 7.8
12:00 236.8 F 12.8
14:00 236.0 F 9.7
16:00 237.2 F 9.2
18:00 246.9 F 14.1
20:00 258.5 F 20.6
22:00 269.2 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.5 F 29.6
2:00 271.9 F 28.8
3:00 271.8 F 29.3
4:00 272.3 F 29.5

Min 271.2 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.4 F 0.0
5:00 214.3 F 0.0
10:00 187.5 F 0.0
15:00 187.9 F 0.0
20:00 172.6 F 0.0
25:00 182.8 F 0.0
30:00 172.6 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 23:04
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:01:36
CYCLE COMPLETE

07/08/2022 02:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____