
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001300
8:52 17/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:47
PRESSURIZING 4:48
STERILIZING 6:06
136.0 C 220kPa 6:06
136.4 C 221kPa 6:36
136.3 C 223kPa 7:06
136.4 C 225kPa 7:36
136.4 C 223kPa 8:06
136.3 C 223kPa 8:36
136.2 C 223kPa 9:06
136.4 C 225kPa 9:36
136.4 C 225kPa 9:36
VENTING 9:36
AIR DRYING 10:35
CYCLE COMPLETE 25:35

Digital Signature #
8484B1933003A570
