
Midmark M11 - v1.0.0
Total Cycles: 5462

Name: RITTER 1
Sterilizer ID: V2178198
03/31/2025 08:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.3 F 0.0
2:00 75.4 F 0.4
4:00 89.4 F 1.9
6:00 123.1 F 4.4
8:00 146.7 F 7.2
10:00 178.4 F 10.7
12:00 205.3 F 2.7
14:00 209.2 F 5.0
16:00 220.6 F 8.2
18:00 229.7 F 11.8
20:00 230.0 F 7.1
22:00 241.8 F 12.1
24:00 237.4 F 9.5
26:00 250.0 F 15.6
28:00 260.7 F 21.5
30:00 270.6 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.2 F 29.2
2:00 271.3 F 28.5
3:00 272.2 F 29.4
4:00 271.7 F 28.6

Min 271.2 F 28.4
Max 272.3 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.9 F 0.0
5:00 184.6 F 0.0
10:00 183.4 F 0.0
15:00 183.7 F 0.0
20:00 183.4 F 0.0
25:00 203.7 F 0.0
30:00 214.1 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:50
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:09:08
CYCLE COMPLETE

03/31/2025 09:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____